

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
04-Jan-16	08:10	5.1	NTH	21:00	4.8	NTH	5.5	4.7	NTH	
05-Jan-16	09:25	5.2	NTH	18:05	5.1	STW	6.4	4.7	NTH	
06-Jan-16	09:00	5.0	NTH	17:05	4.7	STW	5.2	4.6	NTH	
07-Jan-16	09:50	4.7	NTH	17:30	5.1	NTH	5.3	4.5	NTH	
08-Jan-16	08:30	5.0	NTH	17:55	4.9	NTH	5.5	4.6	NTH	
11-Jan-16	10:00	4.9	NTH	17:45	4.8	NTH	5.4	4.7	NTH	
12-Jan-16	09:00	4.7	NTH	17:25	4.9	NTH	5.5	4.6	NTH	
13-Jan-16	08:00	4.8	NTH	17:00	4.8	STW	5.3	4.8	NTH	
14-Jan-16	09:40	5.5	NTH	17:00	5.2	STW	6.1	5.1	NTH	
15-Jan-16	09:15	5.3	NTH	17:20	5.2	NTH	6.3	5.1	NTH	
18-Jan-16	09:00	5.3	NTH	17:20	5.3	STW	5.9	5.2	NTH	
19-Jan-16	09:05	5.2	NTH	17:05	5.6	STW	6.0	5.1	NTH	
20-Jan-16	08:45	5.4	NTH	17:15	5.8	STW	5.9	5.3	NTH	
21-Jan-16	08:55	5.8	NTH	17:15	5.3	STW	6.0	5.3	NTH	
22-Jan-16	09:35	5.7	NTH	17:00	5.7	STW	5.9	5.2	NTH	
25-Jan-16	09:35	4.8	NTH	17:00	4.7	STW	5.4	4.7	NTH	
26-Jan-16	09:15	5.3	NTH	17:20	5.3	STW	5.7	4.9	NTH	
27-Jan-16	09:07	5.7	NTH	17:05	5.7	STW	5.9	5.2	NTH	
28-Jan-16	09:30	5.3	NTH	17:30	5.7	STW	5.9	5.2	NTH	
29-Jan-16	09:25	5.8	NTH	17:30	4.7	NTH	6.5	4.7	NTH	

Nan-Ny 01 Feb 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Current (AM)				Current (PM)			Max. & Min.			Comments
Date (dd-mmm-yy)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Feb-16	08:50	5.0	NTH	17:20	6.1	STW	6.4	4.7	NTH	
02-Feb-16	08:24	5.9	NTH	17:05	5.1	STW	6.4	5.9	NTH	
03-Feb-16	09:20	6.2	NTH	17:10	6.1	STW	6.3	5.9	NTH	
04-Feb-16	08:27	5.9	NTH	17:20	5.9	STW	6.2	5.8	NTH	
05-Feb-16	08:27	5.9	NTH	17:25	6.1	STW	6.3	5.8	NTH	
12-Feb-16	08:17	6.1	NTH	17:10	4.5	NTH	6.4	3.8	NTH	
15-Feb-16	09:06	3.9	NTH	17:00	4.3	STW	4.9	3.8	NTH	
16-Feb-16	09:30	4.7	NTH	17:00	4.5	STW	4.9	3.9	NTH	
17-Feb-16	09:21	4.9	NTH	17:40	4.7	STW	4.9	3.8	NTH	
18-Feb-16	09:36	4.4	NTH	17:10	4.1	STW	4.8	3.8	NTH	
19-Feb-16	09:12	4.8	NTH	17:50	4.1	STW	4.9	3.7	NTH	
22-Feb-16	09:00	4.5	NTH	17:05	4.2	STW	4.9	3.7	NTH	
23-Feb-16	09:30	4.4	NTH	17:05	4.3	STW	5.7	4.1	NTH	
24-Feb-16	08:30	4.4	NTH	17:25	4.3	STW	4.9	3.7	NTH	
25-Feb-16	09:30	3.9	NTH	17:00	4.2	STW	4.9	3.8	NTH	
26-Feb-16	09:44	4.0	NTH	17:17	3.9	NTH	5.1	2.4	NTH	

Valid by 02 Mar 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Mar-16	08:26	3.3	NTH	17:05	3.6	STW	5.3	2.5	NTH	
02-Mar-16	09:40	4.8	NTH	17:25	4.3	STW	4.9	2.8	NTH	
03-Mar-16	09:20	3.1	NTH	17:25	5.7	STW	5.8	3.1	NTH	
04-Mar-16	09:05	5.2	STW	17:05	5.6	NTH	5.8	5.1	NTH	
07-Mar-16	08:35	5.8	NTH	17:00	4.8	STW	5.9	4.7	NTH	
08-Mar-16	09:20	4.7	NTH	17:05	5.6	STW	5.8	4.7	NTH	
09-Mar-16	09:33	5.6	NTH	17:00	5.2	STW	5.7	4.8	NTH	
10-Mar-16	09:11	5.3	NTH	17:25	5.3	STW	6.1	5.1	NTH	
11-Mar-16	08:37	5.2	NTH	17:15	5.6	NTH	5.7	5.1	NTH	
14-Mar-16	08:05	5.4	STW	17:50	5.3	STW	5.9	5.1	NTH	
15-Mar-16	09:39	5.6	NTH	17:26	5.4	NTH	5.8	5.2	NTH	
16-Mar-16	08:19	5.7	NTH	17:24	5.2	NTH	5.8	4.9	NTH	
17-Mar-16	09:36	4.9	NTH	17:00	5.2	STW	5.6	4.8	NTH	
18-Mar-16	10:02	5.1	NTH	17:22	5.4	NTH	5.6	4.9	NTH	
21-Mar-16	08:25	5.1	NTH	17:19	5.3	NTH	5.4	4.7	NTH	
22-Mar-16	08:25 08:25 NTH 22-Mar-16	5.2	NTH	17:05	4.9	STW	5.8	4.7	NTH	
23-Mar-16	08:40	4.7	NTH	17:00	4.3	STW	5.4	4.1	NTH	
24-Mar-16	09:14	4.3	NTH	17:00	4.4	STW	4.9	4.1	NTH	
25-Mar-16	09:01	4.6	NTH	17:22	4.1	NTH	4.9	3.9	NTH	
28-Mar-16	08:25	4.6	NTH	17:05	4.6	STW	4.8	4.1	NTH	
29-Mar-16	09:10	4.6	NTH	17:05	4.3	STW	5.1	4.0	NTH	
30-Mar-16	09:16	4.6	NTH	17:10	4.3	STW	5.1	4.2	NTH	
31-Mar-16	09:35	4.4	NTH	17:00	4.3	STW	5.1	4.0	NTH	

Nair-N by 06Apr2016

Freezer Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Current (AM)				Current (PM)			Max. & Min.			Comments
Date (dd-mmm-yy)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Apr-16	08:20	4.6	NTH	17:05	4.3	STW	4.9	4.1	NTH	
06-Apr-16	08:16	4.7	NTH	17:26	4.3	NTH	4.9	3.9	NTH	
07-Apr-16	09:35	4.4	NTH	17:27	4.5	NTH	4.9	3.9	NTH	
08-Apr-16	08:25	4.4	NTH	17:28	5.2	NTH	5.3	4.0	NTH	
11-Apr-16	08:11	4.8	NTH	17:00	4.6	STW	5.3	4.2	NTH	
12-Apr-16	09:03	4.6	NTH	17:30	4.3	NTH	4.9	4.2	NTH	
13-Apr-16	09:29	4.5	NTH	17:35	4.6	STW	5.1	3.9	NTH	
14-Apr-16	09:25	4.4	NTH	17:05	4.4	STW	5.1	3.9	NTH	
15-Apr-16	09:09	4.6	NTH	17:18	4.0	NTH	4.8	3.9	NTH	
18-Apr-16	10:59	4.0	NTH	17:00	4.3	STW	4.8	3.9	NTH	
19-Apr-16	09:10	4.7	NTH	17:25	4.7	NTH	4.8	3.9	NTH	
20-Apr-16	09:27	4.1	NTH	17:20	4.2	NTH	4.7	3.8	NTH	
21-Apr-16	09:14	4.3	NTH	17:30	4.4	STW	4.7	3.7	NTH	
22-Apr-16	09:20	4.4	NTH	17:05	4.5	NTH	4.6	3.7	NTH	
25-Apr-16	08:53	4.2	NTH	17:00	4.2	STW	4.7	3.8	NTH	
26-Apr-16	09:10	4.6	NTH	17:00	4.0	STW	4.9	3.7	NTH	
27-Apr-16	09:15	4.6	NTH	17:05	4.2	STW	4.8	3.8	NTH	
28-Apr-16	09:01	4.1	NTH	17:57	4.4	NTH	4.8	3.9	NTH	
29-Apr-16	10:00	4.1	NTH	17:05	4.6	STW	4.7	3.9	NTH	

Nair-NTH 03 May 2016

Fridge
Freezer Temperature Log of VGHTC Clinical Trial Pharmacy

NTH 02 May 2016

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
02-May-16	09:05	4.4	NTH	17:00	4.4	STW	4.9	3.9	NTH	
03-May-16	09:14	4.3	NTH	17:05	4.3	STW	5.1	4.1	NTH	
04-May-16	09:43	4.6	NTH	17:30	4.3	STW	4.9	3.9	NTH	
05-May-16	09:05	4.6	NTH	17:20	4.2	STW	4.7	3.9	NTH	
06-May-16	09:00	4.1	NTH	17:30	4.2	NTH	4.8	3.9	NTH	
09-May-16	08:00	4.6	STW	17:45	4.3	STW	4.9	3.6	NTH	
10-May-16	08:05	4.1	STW	17:20	4.3	STW	4.8	3.6	NTH	
11-May-16	08:10	4.1	STW	17:35	4.3	STW	4.7	3.8	NTH	
12-May-16	08:50	4.6	STW	18:40	3.7	STW	4.8	3.7	NTH	
13-May-16	10:22	4.1	NTH	17:13	4.4	NTH	4.7	3.7	NTH	
16-May-16	09:54	4.2	NTH	17:05	4.3	STW	4.8	3.9	NTH	
17-May-16	09:15	4.2	NTH	17:05	4.2	STW	4.8	3.8	NTH	
18-May-16	08:36	4.7	NTH	17:14	4.4	NTH	5.0	3.9	NTH	
19-May-16	09:45	4.1	NTH	17:15	4.4	NTH	4.9	3.7	NTH	
20-May-16	09:10	4.5	NTH	17:30	4.1	NTH	4.8	3.8	NTH	
23-May-16	09:30	4.6	NTH	17:00	4.3	STW	4.8	3.8	NTH	
24-May-16	09:23	4.6	NTH	17:05	4.1	STW	4.9	3.8	NTH	
25-May-16	09:05	4.2	NTH	17:05	3.9	STW	4.8	3.7	NTH	
26-May-16	09:05	4.2	NTH	17:20	4.2	NTH	4.8	3.8	NTH	
27-May-16	08:07	3.9	NTH	17:06	4.2	NTH	4.8	3.7	NTH	
30-May-16	09:15	4.6	NTH	17:05	4.7	NTH	4.9	3.9	NTH	
31-May-16	09:33	4.3	NTH	17:05	4.5	STW	4.8	3.6	NTH	

NTH 03 Jun 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Jun-16	08:28	4.4	NTH	17:05	4.3	STW	4.8	3.7	NTH	
02-Jun-16	08:08	3.9	NTH	17:37	4.4	NTH	5.3	3.3	NTH	
03-Jun-16	08:27	4.4	NTH	17:10	4.6	NTH	5.2	4.3	NTH	
06-Jun-16	08:01	4.9	NTH	17:17	4.6	STW	5.3	4.3	NTH	
07-Jun-16	10:02	4.7	NTH	17:05	4.9	STW	5.8	4.4	NTH	
08-Jun-16	08:59	4.9	NTH	17:15	5.6	STW	5.7	4.7	NTH	
10-Jun-16	10:07	5.2	NTH	17:00	4.4	STW	6.7	4.2	NTH	
13-Jun-16	08:30	4.4	NTH	17:05	4.4	STW	5.6	4.1	NTH	
14-Jun-16	09:15	4.9	NTH	17:20	2.9	STW	5.4	2.4	NTH	
15-Jun-16	09:50	4.7	NTH	17:50	4.8	STW	5.1	4.2	NTH	
16-Jun-16	09:20	5.1	NTH	17:15	4.8	STW	5.2	4.4	NTH	
17-Jun-16	09:40	4.8	NTH	17:10	4.0	NTH	4.9	3.9	NTH	
20-Jun-16	08:35	4.3	NTH	17:35	5.6	NTH	5.6	3.9	NTH	
21-Jun-16	09:35	4.5	NTH	17:05	5.3	STW	5.4	4.2	NTH	
22-Jun-16	08:35	4.5	NTH	17:30	4.4	NTH	5.3	4.2	NTH	
23-Jun-16	09:45	4.7	NTH	17:05	5.2	STW	5.3	4.2	NTH	
24-Jun-16	08:10	5.2	NTH	17:05	5.2	NTH	5.3	4.7	NTH	
27-Jun-16	09:33	4.9	NTH	17:20	5.0	STW	5.3	4.7	NTH	
28-Jun-16	08:45	5.2	NTH	17:15	5.2	NTH	5.3	4.9	NTH	
29-Jun-16	08:10	5.3	NTH	17:05	5.4	STW	5.8	4.9	NTH	
30-Jun-16	08:05	5.4	NTH	17:05	5.4	NTH	5.6	5.1	NTH	

Naili Hy 11 Jul 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Jul-16	09:10	5.2	NTH	18:05	5.3	STW	5.5	4.8	NTH	
04-Jul-16	09:39	4.9	NTH	19:20	4.9	STW	6.4	4.9	NTH	
05-Jul-16	09:00	5.2	NTH	19:00	5.4	STW	6.0	4.9	NTH	
06-Jul-16	08:20	5.9	NTH	19:30	4.8	NTH	6.0	4.3	NTH	
07-Jul-16	09:45	4.7	NTH	19:30	4.4	NTH	4.8	4.2	NTH	
08-Jul-16								NTH 11 Jul 2016		Typhoon off
11-Jul-16	09:45	4.3	NTH	19:05	4.2	STW	4.7	4.2	NTH	
12-Jul-16	09:00	4.7	NTH	19:05	4.7	NTH	5.3	4.2	NTH	
13-Jul-16	08:55	5.2	NTH	19:40	5.1	NTH	5.3	4.7	NTH	
14-Jul-16	09:35	5.1	NTH	18:00	4.9	STW	5.3	4.7	NTH	
15-Jul-16	08:56	5.2	NTH	19:40	5.0	STW	5.3	4.6	NTH	
18-Jul-16	08:10	4.9	STW	19:25	4.7	STW	5.4	4.6	NTH	
19-Jul-16	08:30	4.8	STW	18:10	5.1	STW	5.3	4.6	NTH	
20-Jul-16	08:43	4.9	NTH	19:35	5.4	STW	5.5	4.6	NTH	
21-Jul-16	09:15	5.2	NTH	19:05	4.9	STW	5.6	4.7	NTH	
22-Jul-16	08:20	5.4	NTH	19:00	4.9	STW	5.5	4.8	NTH	
25-Jul-16	08:20	5.3	NTH	19:00	4.7	STW	5.5	4.7	NTH	
26-Jul-16	09:14	4.8	NTH	19:20	4.9	STW	5.3	4.7	NTH	
27-Jul-16	08:05	5.1	STW	18:00	4.9	STW	5.6	4.7	NTH	
28-Jul-16	08:25	5.1	STW	19:25	5.3	STW	5.7	4.9	NTH	
29-Jul-16	08:55	5.1	STW	19:35	5.5	STW	5.6	4.8	NTH	

Nai-King 02 Aug 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

	Current (AM)			Current (PM)			Max. & Min.			
Date (dd-mmm-yy)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	Comments
01-Aug-16	08:49	5.6	NTH	17:30	5.2	NTH	6.8	5.1	NTH	
02-Aug-16	09:20	5.5	NTH	17:10	5.2	STW	5.6	4.8	NTH	
03-Aug-16	08:20	5.2	NTH	17:00	5.2	STW	5.5	4.9	NTH	
04-Aug-16	09:10	5.1	NTH	17:10	5.2	STW	5.5	4.9	NTH	
05-Aug-16	09:25	5.1	NTH	17:20	4.9	STW	5.4	4.7	NTH	
08-Aug-16	08:20	5.3	NTH	17:25	4.7	STW	5.3	4.6	NTH	
09-Aug-16	09:20	5.3	NTH	17:00	4.8	NTH	5.3	4.5	NTH	
10-Aug-16	09:20	4.9	NTH	17:45	4.6	STW	5.1	4.4	NTH	
11-Aug-16	09:20	4.8	NTH	17:05	4.9	STW	5.3	4.6	NTH	
12-Aug-16	08:35	5.1	NTH	17:15	5.0	STW	5.3	4.6	NTH	
15-Aug-16	09:20	4.9	NTH	17:35	4.6	NTH	5.2	4.5	NTH	
16-Aug-16	09:10	4.7	NTH	17:05	4.8	STW	5.3	4.4	NTH	
17-Aug-16	08:05	4.8	NTH	17:15	4.8	STW	5.4	4.6	NTH	
18-Aug-16	09:10	5.2	NTH	17:05	4.8	STW	5.2	4.6	NTH	
19-Aug-16	09:20	4.7	NTH	17:00	4.9	STW	5.5	4.7	NTH	
22-Aug-16	08:40	4.9	NTH	17:30	4.8	NTH	5.3	4.7	NTH	
23-Aug-16	09:45	4.9	NTH	17:15	5.1	NTH	5.3	4.7	NTH	
24-Aug-16	08:30	5.1	NTH	17:35	5.0	NTH	5.2	4.7	NTH	
25-Aug-16	09:42	4.8	NTH	18:00	5.1	NTH	5.3	4.7	NTH	
26-Aug-16	09:27	5.0	NTH	17:05	4.9	NTH	5.3	4.7	NTH	
29-Aug-16	10:50	5.1	NTH	17:05	4.7	STW	5.3	4.7	NTH	
30-Aug-16	08:45	5.2	NTH	17:00	5.1	STW	5.4	4.7	NTH	
31-Aug-16	08:20	5.1	NTH	17:00	5.1	STW	5.3	4.6	NTH	

Nanthy 02 Sep 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Sep-16	09:30	4.7	NTH	17:45	5.0	NTH	5.3	4.7	NTH	
02-Sep-16	08:05	4.8	STH	17:25	4.7	NTH	6.5	4.7	NTH	
05-Sep-16	09:00	4.8	NTH	17:00	4.7	STH	5.3	4.4	NTH	
06-Sep-16	09:10	5.1	NTH	17:15	5.2	NTH	5.4	4.6	NTH	
07-Sep-16	08:54	4.9	NTH	17:15	4.9	STH	6.0	4.8	NTH	
08-Sep-16	09:30	4.9	NTH	17:00	4.2	STH	5.4	4.7	NTH	
09-Sep-16	08:41	5.3	NTH	17:00	4.9	STH	5.3	4.7	NTH	
12-Sep-16	08:40	4.9	NTH	17:00	4.9	STH	5.6	4.7	NTH	
13-Sep-16	09:30	4.9	NTH	17:15	4.7	NTH	5.3	4.4	NTH	
14-Sep-16	09:27	4.6	NTH	17:00	4.8	STH	5.1	4.3	NTH	
15-Sep-16	09:15	4.6	NTH							Moon Festival
16-Sep-16	09:15	4.6	NTH	17:17	4.6	NTH	5.3	4.4	NTH	
19-Sep-16	09:25	4.6	NTH	17:20	5.1	STH	5.7	4.4	NTH	
20-Sep-16	09:27	5.1	NTH	17:05	4.8	STH	5.6	4.7	NTH	
21-Sep-16	09:34	5.3	NTH	17:08	4.8	NTH	5.3	3.9	NTH	
22-Sep-16	09:50	4.9	NTH	17:32	4.4	NTH	4.9	4.0	NTH	
23-Sep-16	09:15	4.2	NTH	17:33	4.6	NTH	4.9	4.0	NTH	
26-Sep-16	09:20	4.7	NTH	17:05	4.3	STH	4.7	4.0	NTH	
27-Sep-16	09:15	4.3	NTH	13:18	4.3	NTH	4.8	4.1	NTH	Typhoon
28-Sep-16	08:12	4.7	NTH	12:09	4.7	NTH	4.9	4.1	NTH	Typhoon
29-Sep-16	09:22	4.4	NTH	17:05	4.4	STH	4.9	3.9	NTH	
30-Sep-16	09:31	4.7	NTH	17:04	4.1	NTH	4.9	3.9	NTH	

Nai-Li Hy 09 Oct 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Current (AM)				Current (PM)			Max. & Min.			Comments
Date (dd-mmm-yy)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
03-Oct-16	09:05	4.6	NTH	18:11	4.5	NTH	4.7	3.8	NTH	
04-Oct-16	10:00	4.2	NTH	17:47	3.9	NTH	5.1	3.8	NTH	
05-Oct-16	09:36	4.4	NTH	17:33	4.0	NTH	4.6	3.8	NTH	
06-Oct-16	09:35	3.9	NTH	17:00	4.5	NTH	4.7	3.7	NTH	
07-Oct-16	09:15	4.1	NTH	17:20	4.4	STW	4.7	3.8	NTH	
11-Oct-16	08:38	4.7	NTH	17:00	4.1	STW	4.8	3.8	NTH	
12-Oct-16	10:22	4.2	NTH	17:15	4.3	STW	4.8	3.9	NTH	
13-Oct-16	09:26	4.4	NTH	17:10	3.9	STW	4.6	3.9	NTH	
14-Oct-16	10:00	4.2	NTH	18:15	4.4	STW	5.1	3.9	NTH	
17-Oct-16	08:38	4.7	NTH	17:00	4.6	STW	5.1	4.3	NTH	
18-Oct-16	09:37	4.8	NTH	17:00	4.7	STW	5.1	4.3	NTH	
19-Oct-16	08:40	4.8	NTH	17:00	4.5	STW	5.0	4.3	NTH	
20-Oct-16	09:18	4.7	NTH	17:00	4.4	STW	4.9	4.2	NTH	
21-Oct-16	09:40	4.8	NTH	17:08	4.7	STW	4.9	4.2	NTH	
24-Oct-16	08:34	4.5	NTH	17:30	4.6	NTH	5.1	4.3	NTH	
25-Oct-16	09:35	4.8	NTH	17:15	4.7	STW	5.0	4.3	NTH	
26-Oct-16	09:43	4.9	NTH	17:00	4.5	STW	5.0	4.3	NTH	
27-Oct-16	09:30	4.6	NTH	17:15	4.3	NTH	5.0	4.2	NTH	
28-Oct-16	09:30	4.8	NTH	17:25	4.3	STW	4.9	4.2	NTH	
31 Oct 16	08:30	4.6	NTH	17:00	4.2	STW	5.1	4.2	NTH	

Nair NTH 02 Nov 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Thermograph: ESCORT, 1031-0078										
	Current (AM)			Current (PM)			Max. & Min.			
Date (dd-mmm-yy)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	Comments
01-Nov-16	09:20	4.5	NTN	17:00	4.7	STW	4.9	4.0	NTN	
02-Nov-16	09:58	4.6	NTN	17:15	4.6	NTN	4.8	4.1	NTN	
03-Nov-16	09:15	4.7	NTN	17:25	4.8	STW	4.8	4.2	NTN	
04-Nov-16	09:40	4.4	NTN	17:00	4.2	STW	4.8	4.1	NTN	
07-Nov-16	09:30	4.8	NTN	17:00	4.7	STW	4.8	4.1	NTN	
08-Nov-16	09:30	4.4	NTN	17:05	4.2	STW	4.8	3.6	NTN	
09-Nov-16	09:30	4.3	NTN	17:00	4.3	STW	4.5	3.4	NTN	
10-Nov-16	09:31	4.4	NTN	17:05	4.3	STW	4.5	3.4	NTN	
11-Nov-16	09:10	4.4	NTN	17:00	4.0	STW	4.5	3.4	NTN	
14-Nov-16	09:50	3.9	NTN	17:05	3.7	STW	4.9	3.3	NTN	
15-Nov-16	09:27	4.2	NTN	17:05	3.5	STW	4.5	3.5	NTN	
16-Nov-16	08:50	3.8	NTN	17:00	3.6	STW	4.5	3.4	NTN	
17-Nov-16	09:18	3.5	NTN	17:10	3.8	STW	4.4	3.5	NTN	
18-Nov-16	09:57	3.8	NTN	17:45	4.4	STW	4.6	3.5	NTN	
21-Nov-16	09:45	4.0	NTN	17:35	4.4	NTN	4.7	3.7	NTN	
22-Nov-16	09:12	3.9	NTN	17:00	5.3	STW	5.6	3.8	NTN	
23-Nov-16	09:45	4.9	NTN	17:20	4.9	STW	5.8	4.8	NTN	
24-Nov-16	10:00	4.9	NTN	17:50	4.9	NTN	5.6	4.8	NTN	
25-Nov-16	09:50	5.3	NTN	17:30	5.3	NTN	5.4	4.8	NTN	
28-Nov-16	09:13	5.3	NTN	17:00	5.4	STW	5.6	4.8	NTN	
29-Nov-16	09:20	5.0	NTN	17:05	5.2	STW	5.6	4.7	NTN	
30-Nov-16	09:50	4.8	NTN	17:05	4.5	STW	5.3	4.3	NTN	

NTN 06 Dec 2016

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Dec-16	09:30	4.6	NTH	17:15	4.4	NTH	5.5	4.2	NTH	
02-Dec-16	09:45	4.3	NTH	17:05	4.7	STH	5.2	4.3	NTH	
05-Dec-16	09:55	4.7	NTH	17:05	4.6	STH	5.9	4.1	NTH	
06-Dec-16	09:25	4.4	NTH	17:05	4.3	STH	4.9	4.3	NTH	
07-Dec-16	09:35	4.5	NTH	17:36	4.6	STH	4.9	4.3	NTH	
08-Dec-16	09:45	4.7	NTH	17:40	4.3	STH	4.9	4.3	NTH	
09-Dec-16	09:30	4.4	NTH	17:10	4.3	STH	4.9	4.2	NTH	
12-Dec-16	08:50	4.5	NTH	18:05	4.6	NTH	4.9	4.2	NTH	
13-Dec-16	09:58	4.4	NTH	18:50	4.7	NTH	5.3	4.1	NTH	
14-Dec-16	10:05	4.8	NTH	17:50	5.1	NTH	5.5	4.7	NTH	
15-Dec-16	09:20	4.8	NTH	17:36	4.9	NTH	5.4	4.4	NTH	
16-Dec-16	09:08	4.9	NTH	17:10	4.9	NTH	5.6	4.7	NTH	
19-Dec-16	08:05	5.0	STH	17:40	4.8	STH	5.5	4.7	NTH	
20-Dec-16	08:05	5.1	STH	18:40	5.2	STH	5.6	4.7	NTH	
21-Dec-16	08:05	5.1	STH	19:05	5.1	STH	5.6	4.7	NTH	
22-Dec-16	08:05	5.1	STH	17:35	5.4	STH	5.6	4.8	NTH	
23-Dec-16	08:00	5.2	STH	17:35	5.3	STH	5.7	4.9	NTH	
26-Dec-16	08:25	5.4	NTH	17:40	5.3	STH	6.8	4.9	NTH	
27-Dec-16	09:40	5.3	NTH	17:00	5.3	STH	5.4	4.9	NTH	
28-Dec-16	09:40	5.0	NTH	17:15	4.9	NTH	5.6	4.9	NTH	
29-Dec-16	09:20	5.1	NTH	17:00	4.7	STH	5.7	4.7	NTH	
30-Dec-16	09:18	5.0	NTH	17:40	4.9	STH	5.4	4.7	NTH	

NTH 03 Jan 2017