


TCVGH Temperature Log



| | |
|---------------|-----------------------|
| Sensor Name | Internal |
| Legend | |
| Description | PHCL-C009 |
| Serial Number | 1931-0084 |
| Product Code | EI-IN-D-32-L |
| Start | 2026/3/31 下午 05:00:00 |
| Finish | 2026/5/4 上午 08:20:00 |
| Sensor Range | -40.5 to 70.4 °C |
| Specification | 2.0 to 8.0 °C |
| Average | 5.2 °C |
| Highest | 6.2 °C |
| Lowest | 4.3 °C |

Yuanasy
05 May 2026