


# TCVGH Temperature Log

Internal (°C)



2025/12/31 下午 04:31:00    2026/1/5 下午 04:31:00    2026/1/10 下午 04:31:00    2026/1/15 下午 04:31:00    2026/1/20 下午 04:31:00    2026/1/25 下午 04:31:00    2026/1/30 下午 04:31:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2025/12/31 下午 04:31:00
Finish	2026/2/2 上午 08:11:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.0 °C
Highest	5.8 °C
Lowest	4.4 °C

*tuansy*  
*02 Feb 2026*