


TCVGH Temperature Log

Internal (°C)



2026/7/31 下午 01:00:00 2026/8/5 下午 01:00:00 2026/8/10 下午 01:00:00 2026/8/15 下午 01:00:00 2026/8/20 下午 01:00:00 2026/8/25 下午 01:00:00 2026/8/30 下午 01:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C019
Serial Number	2247-0167
Product Code	EI-IN-D-32-L
Start	2026/7/31 下午 01:00:00
Finish	2026/9/1 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.1 °C
Highest	5.5 °C
Lowest	4.7 °C

Yuan-shu
01 Sep 2026