


TCVGH Temperature Log

Internal (°C)



Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2025/10/31 下午 05:00:00
Finish	2025/12/1 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.9 °C
Highest	5.8 °C
Lowest	3.9 °C

Quarant
01 Dec 2025