


TCVGH Temperature Log

Internal (°C)



2025/2/27 下午 04:00:00 2025/3/4 下午 04:00:00 2025/3/9 下午 04:00:00 2025/3/14 下午 04:00:00 2025/3/19 下午 04:00:00 2025/3/24 下午 04:00:00 2025/3/29 下午 04:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C015
Serial Number	1803-0051
Product Code	EI-IN-D-32-L
Start	2025/2/27 下午 04:00:00
Finish	2025/4/1 上午 08:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.3 °C
Highest	6.2 °C
Lowest	4.6 °C

YucnerSu
01 Apr 2025