


TCVGH Temperature Log

Internal (°C)



2024/10/30 下午 08:00:00 2024/11/4 下午 08:00:00 2024/11/9 下午 08:00:00 2024/11/14 下午 08:00:00 2024/11/19 下午 08:00:00 2024/11/24 下午 08:00:00 2024/11/29 下午 08:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C009
Serial Number	1931-0087
Product Code	EI-IN-D-32-L
Start	2024/10/30 下午 08:00:00
Finish	2024/12/2 下午 05:00:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.3 °C
Highest	6.4 °C
Lowest	4.5 °C

Yuaner Sun
03 Dec 2024