


TCVGH Temperature Log

Internal (°C)



2024/11/29 下午 05:00:00 2024/12/4 下午 05:00:00 2024/12/9 下午 05:00:00 2024/12/14 下午 05:00:00 2024/12/19 下午 05:00:00 2024/12/24 下午 05:00:00 2024/12/29 下午 05:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1715-0143
Product Code	EI-IN-D-32-L
Start	2024/11/29 下午 05:00:00
Finish	2025/1/2 上午 09:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.3 °C
Highest	6.6 °C
Lowest	4.2 °C

Yuaner Su
20 Jan 2025