


# TCVGH Temperature Log

Internal (°C)



2026/1/30 下午 05:00:00    2026/2/4 下午 05:00:00    2026/2/9 下午 05:00:00    2026/2/14 下午 05:00:00    2026/2/19 下午 05:00:00    2026/2/24 下午 05:00:00    2026/3/1 下午 05:00:00

|               |                       |
|---------------|-----------------------|
| Sensor Name   | Internal              |
| Legend        |                       |
| Description   | PHCL-C016             |
| Serial Number | 1715-0134             |
| Product Code  | EI-IN-D-32-L          |
| Start         | 2026/1/30 下午 05:00:00 |
| Finish        | 2026/3/2 上午 08:20:00  |
| Sensor Range  | -40.5 to 70.4 °C      |
| Specification | 2.0 to 8.0 °C         |
| Average       | 5.2 °C                |
| Highest       | 6.6 °C                |
| Lowest        | 4.1 °C                |

*Yuan Su*  
*02 Mar 2026*