


TCVGH Temperature Log

Internal (°C)



2022/9/30 下午 01:30:00 2022/10/5 下午 01:30:00 2022/10/10 下午 01:30:00 2022/10/15 下午 01:30:00 2022/10/20 下午 01:30:00 2022/10/25 下午 01:30:00 2022/10/30 下午 01:30:00

Sensor Name	Internal
Legend	
Description	PHCL-C006
Serial Number	1715-0134
Product Code	EI-IN-D-32-L
Start	2022/9/30 下午 01:30:00
Finish	2022/11/1 上午 09:00:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.4 °C
Highest	6.1 °C
Lowest	4.5 °C

Yuchen Su

01 NOV 2022