


TCVGH Temperature Log

Internal (°C)



2025/4/30 上午 11:00:00 2025/5/5 上午 11:00:00 2025/5/10 上午 11:00:00 2025/5/15 上午 11:00:00 2025/5/20 上午 11:00:00 2025/5/25 上午 11:00:00 2025/5/30 上午 11:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C015
Serial Number	1803-0051
Product Code	EI-IN-D-32-L
Start	2025/4/30 上午 11:00:00
Finish	2025/6/2 上午 10:40:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.9 °C
Highest	5.7 °C
Lowest	4.1 °C

YuanenSu
02 Jun 2025