


TCVGH Temperature Log

Internal (°C)



2024/11/29 下午 05:30:00 2024/12/4 下午 05:30:00 2024/12/9 下午 05:30:00 2024/12/14 下午 05:30:00 2024/12/19 下午 05:30:00 2024/12/24 下午 05:30:00 2024/12/29 下午 05:30:00

Sensor Name	Internal
Legend	
Description	PHCL-C018
Serial Number	2247-0162
Product Code	EI-IN-D-32-L
Start	2024/11/29 下午 05:30:00
Finish	2025/1/2 上午 09:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.5 °C
Highest	6.6 °C
Lowest	3.3 °C

Yuaner Su

02 Jan 2025