

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C003 (Model: R-1500)

Thermograph: ESCORT,1116-0018

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Mar-17	09:10	4.6	NTN	17:15	3.6	STW	5.2	3.0	NTN	
02-Mar-17	09:24	3.7	NTN	17:22	3.4	NTN	5.5	2.7	NTN	
03-Mar-17	09:28	5.1	NTN	17:00	5.1	STW	5.5	4.9	NTN	
06-Mar-17	08:25	5.1	NTN	17:00	5.6	STW	5.6	4.7	NTN	
07-Mar-17	09:13	5.2	NTN	17:10	5.1	STW	5.7	4.9	NTN	
08-Mar-17	09:30	5.2	NTN	17:05	5.1	NTN	5.6	4.6	NTN	
09-Mar-17	09:20	4.8	NTN	17:45	4.2	STW	5.4	4.2	NTN	
10-Mar-17	08:14	5.1	NTN	17:30	4.9	STW	5.6	4.7	NTN	
13-Mar-17	08:05	5.0	STW	17:35	5.4	STW	5.7	4.8	NTN	
14-Mar-17	08:05	4.9	STW	17:25	5.1	STW	5.7	4.8	NTN	
15-Mar-17	08:05	4.9	STW	17:40	5.4	STW	5.7	4.7	NTN	
16-Mar-17	08:05	5.6	STW	17:25	5.1	STW	5.7	4.8	NTN	
17-Mar-17	08:05	4.9	STW	17:20	5.1	STW	5.7	4.8	NTN	
20-Mar-17	08:40	5.0	NTN	17:05	4.9	STW	5.7	4.8	NTN	
21-Mar-17	09:30	5.1	NTN	17:45	5.1	NTN	5.7	4.7	NTN	
22-Mar-17	08:15	5.2	NTN	18:15	4.8	STW	5.6	4.6	NTN	
23-Mar-17	08:15	4.9	NTN	17:00	4.9	STW	5.5	4.7	NTN	
24-Mar-17	10:00	4.9	NTN	17:05	5.0	STW	5.5	4.6	NTN	
27-Mar-17	08:40	4.7	NTN	17:05	4.7	STW	5.7	4.6	NTN	
28-Mar-17	09:30	4.8	NTN	16:55	5.2	STW	5.5	4.4	NTN	
29-Mar-17	09:30	4.9	NTN	17:30	4.7	STW	5.5	4.5	NTN	
30-Mar-17	09:10	4.7	NTN	17:05	5.2	STW	5.3	4.3	NTN	
31-Mar-17	10:05	4.7	NTN	17:15	4.6	NTN	5.4	4.5	NTN	

Nai-Ky 05 April 2017