


TCVGH Temperature Log

Internal (°C)



2025/7/31 下午 05:00:00 2025/8/5 下午 05:00:00 2025/8/10 下午 05:00:00 2025/8/15 下午 05:00:00 2025/8/20 下午 05:00:00 2025/8/25 下午 05:00:00 2025/8/30 下午 05:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1715-0143
Product Code	EI-IN-D-32-L
Start	2025/7/31 下午 05:00:00
Finish	2025/9/1 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.8 °C
Highest	6.6 °C
Lowest	3.9 °C

YuanenSu
01 Sep 2025