


TCVGH Temperature Log

Internal (°C)



2025/11/28 下午 12:00:00 2025/12/3 下午 12:00:00 2025/12/8 下午 12:00:00 2025/12/13 下午 12:00:00 2025/12/18 下午 12:00:00 2025/12/23 下午 12:00:00 2025/12/28 下午 12:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C008
Serial Number	1931-0074
Product Code	EI-IN-D-32-L
Start	2025/11/28 下午 12:00:00
Finish	2026/1/2 上午 08:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.3 °C
Highest	6.1 °C
Lowest	4.6 °C

Yucen Su

02 Jan 2016