


TCVGH Temperature Log

Internal (°C)



2024/12/31 下午 12:00:00 2025/1/5 下午 12:00:00 2025/1/10 下午 12:00:00 2025/1/15 下午 12:00:00 2025/1/20 下午 12:00:00 2025/1/25 下午 12:00:00 2025/1/30 下午 12:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2024/12/31 下午 12:00:00
Finish	2025/2/3 下午 01:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.5 °C
Highest	6.2 °C
Lowest	4.4 °C

Yuenen

03 Feb 2025