


TCVGH Temperature Log

Internal (°C)



2026/5/29 下午 03:00:00	2026/6/3 下午 03:00:00	2026/6/8 下午 03:00:00	2026/6/13 下午 03:00:00	2026/6/18 下午 03:00:00	2026/6/23 下午 03:00:00	2026/6/28 下午 03:00:00
Sensor Name	Internal					
Legend						
Description	PHCL-C009					
Serial Number	1931-0084					
Product Code	EI-IN-D-32-L					
Start	2026/5/29 下午 03:00:00					
Finish	2026/7/1 上午 08:30:00					
Sensor Range	-40.5 to 70.4 °C					
Specification	2.0 to 8.0 °C					
Average	5.1 °C					
Highest	6.1 °C					
Lowest	4.5 °C					

Yuanesin
01 Jul 2026