


TCVGH Temperature Log

Internal (°C)



Sensor Name	Internal
Legend	
Description	PHCL-C001
Serial Number	1715-0141
Product Code	EI-IN-D-32-L
Start	2019/1/31 下午 06:20:00
Finish	2019/3/1 上午 11:00:00
Specification	2.0 to 8.0 °C
Average	5.1 °C
Highest	6.5 °C
Lowest	3.9 °C

Na-z ky 01 Mar 2019