


TCVGH Temperature Log

Internal (°C)



2025/8/29 下午 05:00:00	2025/9/3 下午 05:00:00	2025/9/8 下午 05:00:00	2025/9/13 下午 05:00:00	2025/9/18 下午 05:00:00	2025/9/23 下午 05:00:00	2025/9/28 下午 05:00:00
Sensor Name	Internal					
Legend						
Description	PHCL-C012					
Serial Number	1803-0047					
Product Code	EI-IN-D-32-L					
Start	2025/8/29 下午 05:00:00					
Finish	2025/10/1 下午 01:20:00					
Sensor Range	-40.5 to 70.4 °C					
Specification	2.0 to 8.0 °C					
Average	4.8 °C					
Highest	5.7 °C					
Lowest	4.2 °C					

Yuaner Su

01 Oct 2025