


TCVGH Room Temperature Log

Air Temp (°C)



2026/5/28 下午 09:00:00 2026/6/2 下午 09:00:00 2026/6/7 下午 09:00:00 2026/6/12 下午 09:00:00 2026/6/17 下午 09:00:00 2026/6/22 下午 09:00:00 2026/6/27 下午 09:00:00

Sensor Name	Air Temp
Legend	
Description	PHCL-C021(室)
Serial Number	1333-0017
Product Code	EI-HS-D-32-L
Start	2026/5/28 下午 09:00:00
Finish	2026/7/1 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	15.0 to 25.0 °C
Average	20.8 °C
Highest	21.3 °C
Lowest	20.1 °C

Yuan
01 Jul 2026