


TCVGH Temperature Log

Internal (°C)



2024/8/30 下午 05:00:00 2024/9/4 下午 05:00:00 2024/9/9 下午 05:00:00 2024/9/14 下午 05:00:00 2024/9/19 下午 05:00:00 2024/9/24 下午 05:00:00 2024/9/29 下午 05:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2024/8/30 下午 05:00:00
Finish	2024/10/1 下午 01:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.8 °C
Highest	5.7 °C
Lowest	4.1 °C

YuanenSu
01 Oct 2024