


# TCVGH Temperature Log

Internal (°C)



2025/2/27 下午 04:00:00    2025/3/4 下午 04:00:00    2025/3/9 下午 04:00:00    2025/3/14 下午 04:00:00    2025/3/19 下午 04:00:00    2025/3/24 下午 04:00:00    2025/3/29 下午 04:00:00

|               |                       |
|---------------|-----------------------|
| Sensor Name   | Internal              |
| Legend        |                       |
| Description   | PHCL-C019             |
| Serial Number | 2247-0060             |
| Product Code  | EI-IN-D-32-L          |
| Start         | 2025/2/27 下午 04:00:00 |
| Finish        | 2025/4/1 上午 09:50:00  |
| Sensor Range  | -40.5 to 70.4 °C      |
| Specification | 2.0 to 8.0 °C         |
| Average       | 5.3 °C                |
| Highest       | 6.6 °C                |
| Lowest        | 4.7 °C                |

*Yucnelisa*  
*01 Apr 2025*