

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C003 (Model: R-1500)

Thermograph: ESCORT,1116-0018 (from 01Jun2017 to 05Jun2017) : 1043-0065 (05Jun2017 - 30Jun2017)

	Current (AM)			Current (PM)			Max. & Min.			
Date (dd-mmm-yy)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	Comments
01-Jun-17	08:45	4.2	~YH	17:38	4.6	~YH	4.9	3.9	~YH	
02-Jun-17	08:24	4.4	~YH	17:12	4.2	STW	4.7	3.9	~YH	
05-Jun-17	08:35	4.0	~YH	17:00	4.3	STW	5.4	3.9	~YH	1043-006
06-Jun-17	08:05	4.2	~YH	17:00	3.2	STW	4.9	2.6	~YH	
07-Jun-17	08:15	3.7	~YH	17:25	4.6 4.6	STW	4.8	3.3	~YH	
08-Jun-17	08:44	4.3	~YH	17:43	4.3	~YH	4.8	3.8	~YH	
09-Jun-17	08:15	3.9	~YH	17:00	4.0	~YH	4.6	3.6	~YH	
12-Jun-17	08:20	4.5	~YH	17:00	3.9	STW	4.6	3.6	~YH	
13-Jun-17	08:16	4.1	~YH	17:35	3.9	STW	4.7	3.8	~YH	
14-Jun-17	08:35	4.2	~YH	17:10	4.2	STW	4.8	3.7	~YH	
15-Jun-17	08:14	4.1	~YH	17:00	4.4	STW	5.1	3.9	~YH	
16-Jun-17	08:38	4.6	~YH	17:00	4.4	STW	5.0	3.9	~YH	
19-Jun-17	08:14	4.5	~YH	17:05	4.5	STW	5.1	4.3	~YH	
20-Jun-17	08:45	4.7	~YH	17:25	4.5	~YH	5.0	3.7	~YH	
21-Jun-17	08:35	3.7	~YH	17:00	4.1	STW	4.8	3.2	~YH	
22-Jun-17	08:15	4.6	~YH	17:00	4.3	STW	4.8	3.9	~YH	
23-Jun-17	08:55	4.3	~YH	17:20	3.9	STW	4.9	3.6	~YH	
26-Jun-17	08:15	4.1	~YH	17:25	4.1	~YH	4.5	3.5	~YH	
27-Jun-17	08:15	3.9	~YH	17:35	4.2	~YH	4.6	3.3	~YH	
28-Jun-17	08:25	4.2	~YH	17:00	4.2	STW	4.9	3.7	~YH	
29-Jun-17	08:10	4.4	~YH	17:00	4.2	STW	4.9	3.7	~YH	
30-Jun-17	08:48	4.4	~YH	17:35	4.3	~YH	5.1	4.0	~YH	

Na-TH 03Jul 2017