


TCVGH Temperature log



|               |                       |
|---------------|-----------------------|
| Sensor Name   | Internal              |
| Legend        |                       |
| Description   | PHCL-C018             |
| Serial Number | 2247-0162             |
| Product Code  | EI-IN-D-32-L          |
| Start         | 2025/3/31 下午 05:00:00 |
| Finish        | 2025/5/1 上午 09:50:00  |
| Sensor Range  | -40.5 to 70.4 °C      |
| Specification | 2.0 to 8.0 °C         |
| Average       | 4.9 °C                |
| Highest       | 6.6 °C                |
| Lowest        | 4.4 °C                |

Unwenwang  
01 May 2025