


TCVGH Temperature log



Sensor Name	Internal
Legend	
Description	PHCL-C012
Serial Number	1715-0141
Product Code	EI-IN-D-32-L
Start	2025/3/31 下午 05:00:00
Finish	2025/5/1 上午 09:40:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.6 °C
Highest	6.7 °C
Lowest	4.7 °C

Unwennang
01 May 2025