


TCVGH Temperature Log

Internal (°C)



2026/4/30 下午 05:00:00 2026/5/5 下午 05:00:00 2026/5/10 下午 05:00:00 2026/5/15 下午 05:00:00 2026/5/20 下午 05:00:00 2026/5/25 下午 05:00:00 2026/5/30 下午 05:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2026/4/30 下午 05:00:00
Finish	2026/6/1 下午 02:00:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.8 °C
Highest	5.9 °C
Lowest	4.0 °C

Yuanes
01 Jun 2026