


TCVGH Temperature Log

Internal (°C)



2026/3/31 下午 05:00:00 2026/4/5 下午 05:00:00 2026/4/10 下午 05:00:00 2026/4/15 下午 05:00:00 2026/4/20 下午 05:00:00 2026/4/25 下午 05:00:00 2026/4/30 下午 05:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C018
Serial Number	2247-0162
Product Code	EI-IN-D-32-L
Start	2026/3/31 下午 05:00:00
Finish	2026/5/4 上午 08:10:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.6 °C
Highest	5.1 °C
Lowest	4.3 °C

Yuanan
05 May 2026