


TCVGH Temperature Log

Internal (°C)



2026/5/29 下午 03:00:00 2026/6/3 下午 03:00:00 2026/6/8 下午 03:00:00 2026/6/13 下午 03:00:00 2026/6/18 下午 03:00:00 2026/6/23 下午 03:00:00 2026/6/28 下午 03:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C008
Serial Number	1931-0074
Product Code	EI-IN-D-32-L
Start	2026/5/29 下午 03:00:00
Finish	2026/7/1 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.4 °C
Highest	6.9 °C
Lowest	4.6 °C

Yuanesun
01/Jul 2026