


TCVGH Temperature Log

Internal (°C)



Sensor Name	Internal
Legend	
Description	PHCL-C001
Serial Number	1715-0141
Product Code	EI-IN-D-32-L
Start	2019/3/29 下午 12:00:00
Finish	2019/5/1 上午 08:30:00
Specification	2.0 to 8.0 °C
Average	5.4 °C
Highest	6.8 °C
Lowest	3.7 °C

Nā-zny 01 May 2019