


TCVGH Temperature log



	2018/5/31 上午 09:15:00	2018/6/5 上午 09:15:00	2018/6/10 上午 09:15:00	2018/6/15 上午 09:15:00	2018/6/20 上午 09:15:00	2018/6/25 上午 09:15:00	2018/6/30 上午 09:15:00
Sensor Name	Internal						

Legend

Description	PHCL-C001
Serial Number	1715-0141
Product Code	EI-IN-D-32-L
Start	2018/5/31 上午 09:15:00
Finish	2018/7/2 上午 08:25:00
Specification	2.0 to 8.0 °C
Average	4.9 °C
Highest	7.1 °C
Lowest	2.7 °C

Yu Wen Wang 02 Jul 2018