


TCVGH Temperature log



PHCL-C001

Sensor Name	Internal
Legend	
Description	PHCL-C001
Serial Number	1715-0141
Product Code	EI-IN-D-32-L
Start	2018/1/31 下午 01:00:00
Finish	2018/3/1 上午 09:10:00
Specification	2.0 to 8.0 °C
Average	4.7 °C
Highest	7.2 °C
Lowest	3.1 °C

Nā-z ny 01 Mar 2018