


TCVGH Temperature Log

Internal (°C)



2024/5/31 下午 01:00:00	2024/6/5 下午 01:00:00	2024/6/10 下午 01:00:00	2024/6/15 下午 01:00:00	2024/6/20 下午 01:00:00	2024/6/25 下午 01:00:00	2024/6/30 下午 01:00:00
--------------------------	-------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------

Sensor Name	Internal
Legend	
Description	PHCL-C008
Serial Number	1931-0074
Product Code	EI-IN-D-32-L
Start	2024/5/31 下午 01:00:00
Finish	2024/7/1 上午 08:40:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.4 °C
Highest	6.3 °C
Lowest	4.7 °C

Yuchen Su
01 Jul 2024