


TCVGH Temperature Log

Internal (°C)



Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1715-0143
Product Code	EI-IN-D-32-L
Start	2026/3/31 下午 05:00:00
Finish	2026/5/4 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.1 °C
Highest	5.9 °C
Lowest	4.4 °C

Uwanesu
05 May 2016