


TCVGH Temperature Log

Internal (°C)



2025/4/30 上午 11:00:00	2025/5/5 上午 11:00:00	2025/5/10 上午 11:00:00	2025/5/15 上午 11:00:00	2025/5/20 上午 11:00:00	2025/5/25 上午 11:00:00	2025/5/30 上午 11:00:00
Sensor Name	Internal					
Legend						
Description	PHCL-C009					
Serial Number	1931-0087					
Product Code	EI-IN-D-32-L					
Start	2025/4/30 上午 11:00:00					
Finish	2025/6/2 上午 10:30:00					
Sensor Range	-40.5 to 70.4 °C					
Specification	2.0 to 8.0 °C					
Average	5.4 °C					
Highest	7.5 °C					
Lowest	4.7 °C					

Yumen Su
02 Jun 2025