

STATEMENT OF ACCURACY

ILog Temperature Data loggers

ILog data loggers are calibrated for temperature in C/T's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter: Fluke - Hart Scientific 1521 Digital Thermometer Readout w/ Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: EI-IN-D-32-L
 Serial Number: 1803-0047
 Date: 23 Jan 2018



Cryopak Verification Technologies
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Rebekah Crawford (Production Supervisor)
 Parminder Singh (Quality Manager)

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008. ACC-031 Revision 2 Page Form Page 1/1.



STATEMENT OF ACCURACY

ILog Temperature Data loggers

ILog data loggers are calibrated for temperature in C/T's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter: Fluke - Hart Scientific 1521 Digital Thermometer Readout w/ Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: EI-IN-D-32-L
 Serial Number: 1803-0048
 Date: 23 Jan 2018



Cryopak Verification Technologies
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Rebekah Crawford (Production Supervisor)
 Parminder Singh (Quality Manager)

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008. ACC-031 Revision 2 Page Form Page 1/1.



STATEMENT OF ACCURACY

ILog Temperature Data loggers

ILog data loggers are calibrated for temperature in C/T's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter: Fluke - Hart Scientific 1521 Digital Thermometer Readout w/ Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.


Product Code: EI-IN-D-32-L
 Serial Number: 1803-0049
 Date: 23 Jan 2018



Cryopak Verification Technologies
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Rebekah Crawford (Production Supervisor)
 Parminder Singh (Quality Manager)

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008. ACC-031 Revision 2 Page Form Page 1/1.



STATEMENT OF ACCURACY

iLog Temperature Data loggers

Log data loggers are calibrated for temperature in CPT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter, Fluke-Hart Scientific 1521 Digital Thermometer Readout w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: **EI-IN-D-32-L**
 Serial Number: **1803-0050**
 Date: **23 Jan 2018**

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
X	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C	± 0.30°C
✓	-40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C		
X	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C	± 0.30°C
X	0°C to +150°C	0°C to +90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C	± 0.75°C

Rebekah Crawford
 Rebekah Crawford (Production Supervisor)

Parminder Singh
 Parminder Singh (Quality Manager)

bsi.
 ISO 9001:2008

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/1.

STATEMENT OF ACCURACY

iLog Temperature Data loggers

Log data loggers are calibrated for temperature in CPT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter, Fluke-Hart Scientific 1521 Digital Thermometer Readout w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: **EI-IN-D-32-L**
 Serial Number: **1803-0051**
 Date: **23 Jan 2018**

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
X	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C	± 0.30°C
✓	-40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C		
X	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C	± 0.30°C
X	0°C to +150°C	0°C to +90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C	± 0.75°C

Rebekah Crawford
 Rebekah Crawford (Production Supervisor)

Parminder Singh
 Parminder Singh (Quality Manager)

bsi.
 ISO 9001:2008

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/1.

