


# TCVGH Room Temperature Log




Sensor Name	Air Temp
Legend	
Description	Room temperature
Serial Number	1333-0016
Product Code	EI-HS-D-32-L
Start	2022/6/1 上午 10:40:00
Finish	2022/7/18 上午 04:50:00
Sensor Range	-40.5 to 70.4 °C
Specification	15.0 to 25.0 °C
Average	20.9 °C
Highest	23.0 °C
Lowest	17.8 °C

Because electronic temperature recorder (serial number: 1333-0016) was out of power on 18 Jul 2022, so providing back up temperature records (serial number: 1333-0017)

Yuchen Su 01 Aug 2022

# TCVGH Room Temperature Log




Sensor Name	Air Temp
Legend	
Description	Room temperature
Serial Number	1333-0016
Product Code	EI-HS-D-32-L
Start	2022/7/18 上午 08:15:00
Finish	2022/8/1 上午 10:35:00
Sensor Range	-40.5 to 70.4 °C
Specification	15.0 to 25.0 °C
Average	20.7 °C
Highest	22.6 °C
Lowest	19.3 °C

YUCHE SU  
01 Aug 2022



# TCVGH Room Temperature Log



Sensor Name	Air Temp
Legend	
Description	room temperature 2
Serial Number	1333-0017
Product Code	EI-HS-D-32-L
Start	2022/7/1 下午 06:00:00
Finish	2022/8/1 上午 10:30:00
Sensor Range	-40.5 to 70.4 °C
Specification	15.0 to 25.0 °C
Average	20.4 °C
Highest	22.4 °C
Lowest	17.9 °C

*Yu Chen Sy*  
*01 Aug 2022*