

STATEMENT OF ACCURACY

IIlog Temperature Data loggers

IIlog data loggers are calibrated for temperature in C17's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter, Fluke - Hart Scientific 1531 Digital Thermometer Readout w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: EI-IN-D-32-L
 Serial Number: 1715-0115
 Date: 17 Aug 2017
Cryopak
 Verification Technologies
 a Toy Company

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
X	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C	± 0.30°C
✓	40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C		
X	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C	± 0.30°C
X	0°C to +150°C	0°C to -90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C	± 0.75°C

Rohit Khandelwal
 Relehan Crawford (production supervisor)

Permlinder Singh (Quality Manager)

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Authorized Quality Document for Cryopak Verification Technologies Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/L.

bsi.
 ISO 9001:2008

STATEMENT OF ACCURACY

IIlog Temperature Data loggers

IIlog data loggers are calibrated for temperature in C17's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST), Temperature Reference Meter, Fluke - Hart Scientific 1531 Digital Thermometer Readout w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: EI-IN-D-32-L
 Serial Number: 1715-0134
 Date: 17 Aug 2017
Cryopak
 Verification Technologies
 a Toy Company

Temp. Code	Full Range	Temp. Range	Accuracy	Temp. Range	Accuracy	Temp. Range	Accuracy
X	-20°C to +100°C	-20°C to 0°C	± 0.30°C	0°C to +70°C	± 0.25°C	+70°C to +100°C	± 0.30°C
✓	40°C to +70°C	-40°C to 0°C	± 0.35°C	0°C to +70°C	± 0.25°C		
X	-100°C to +40°C	-100°C to -80°C	± 1.0°C	-80°C to -50°C	± 0.50°C	-50°C to +40°C	± 0.30°C
X	0°C to +150°C	0°C to -90°C	± 0.30°C	+90°C to +120°C	± 0.45°C	+120°C to +150°C	± 0.75°C

Rohit Khandelwal
 Relehan Crawford (production supervisor)

Permlinder Singh (Quality Manager)

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24066
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Authorized Quality Document for Cryopak Verification Technologies Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/L.

bsi.
 ISO 9001:2008



STATEMENT OF ACCURACY

II-Log Temperature Data loggers

II-Log data loggers are calibrated for temperature in CTT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke - Hart Scientific 1521 Digital Thermometer. Recheck w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: **EI-IN-D-32-L**
 Serial Number: **1715-0141**
 Date: **17 Aug 2017**
Cryopak
 Verification Technologies
 a TCR Company

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24666
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Rebekah Crawford (Production Supervisor)

Farminder Singh (Quality Manager)

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/1.



STATEMENT OF ACCURACY

II-Log Temperature Data loggers

II-Log data loggers are calibrated for temperature in CTT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke - Hart Scientific 1521 Digital Thermometer. Recheck w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: **EI-IN-D-32-L**
 Serial Number: **1715-0142**
 Date: **17 Aug 2017**
Cryopak
 Verification Technologies
 a TCR Company

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24666
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Rebekah Crawford (Production Supervisor)

Farminder Singh (Quality Manager)

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/1.



STATEMENT OF ACCURACY

II-Log Temperature Data loggers

II-Log data loggers are calibrated for temperature in CTT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke - Hart Scientific 1521 Digital Thermometer. Recheck w Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate.

Product Code: **EI-IN-D-32-L**
 Serial Number: **1715-0143**
 Date: **17 Aug 2017**
Cryopak
 Verification Technologies
 a TCR Company

Cryopak Verification Technologies, Inc.
 PO Box 309, 120 Parkway Drive
 Buchanan, VA 24666
 Phone: 540-254-1433 | Fax: 540-254-2433
 techsupport@cryopak.com
 www.cryopak.com

Rebekah Crawford (Production Supervisor)

Farminder Singh (Quality Manager)

Authorized Quality Document for Cryopak Verification Technologies. Reference: ISO 9001:2008, ACC-031 Revision 2 Page Form Page 1/1.

