



溫度部份

器示值 (°C)	標準值 (°C)	器差值 (°C)
2.0	2.3	-0.3
4.4	4.8	-0.4
6.9	7.4	-0.5

校正說明：

1. 器差值 = 器示值 - 標準值
 2. 器差% = $[(\text{器示值} - \text{標準值}) \div \text{標準值}] * 100$
 3. 器示值係指送校正件所顯示或設定之值
 4. 標準值係指工作標準件之輸出值或顯示值
 5. 校正能力係以約 95% 信賴水準, $k=2$ 之擴充不確定度表示
 6. 擴充不確定度: 1.0 °C
 7. 上述擴充不確定度已包含待校件之不確定度評估結果
- THE END --



This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

TWB6688664



溫度部份：

器示值 (°C)	標準值 (°C)	器差值 (°C)
-19.6	-20.0	0.4
2.3	2.0	0.3
8.2	8.0	0.2

校正說明：

1. 器差值 = 器示值 - 標準值
 2. 器差% = $[(\text{器示值} - \text{標準值}) / \text{標準值}] * 100$
 3. 器示值係指送校正件所顯示或設定之值
 4. 標準值係指工作標準件之輸出值或顯示值
 5. 校正能力係以約95%信賴水準, $k=2$ 之擴充不確定度表示
 6. 擴充不確定度: 0.5°C
 7. 上述擴充不確定度已包含待校件之不確定度評估結果
- THE END --



校正報告書

校正暨量測實驗室-台中

		Report No : ETAC1368517	第 1 頁	共 2 頁
申請者 Applicant	臺中榮民總醫院			
儀器名稱 Equipment	溫度蒐集器			
製造廠商 Manufacturer	ESCORT	機型 Model	E1-IN-D-32-L	序號 Serial No
校正程序 Procedure used	GENP-ET-T008	收件日期 Received Date	2017/5/8	校正日期 Calibration Date
校驗者 Operator	陳哲嘉	溫度 °C Temperature	23 °C ± 2 °C	相對濕度 % Relative Humidity
顧客地址 送校地址	臺中市407西屯區臺灣大道四段1650號			校正地點 Location
	實驗室			

實驗室使用標準器 / SGS Standards

儀器名稱 Equipment	製造廠商 Manufacturer	機型 Model	標準器校正日期 Calibration Date
Temp & Humidity Chamber	BINDER	-	2016/11/21
HYGRO PALM SERIE	ROTRONIC	HP22-A	2016/8/18
序號 Serial Number	追溯單位 Traceability	報告號碼 Report No.	標準器有效日期 Due Date
S0006989	SGS(TCH)	ETAC3276116	2017/11/20
61241484	SGS(TAF 0143)	ECTC2505716	2017/8/17

- ◆ 台灣檢驗科技股份有限公司特此聲明本報告書內記載之標準器，依ILAC P10之規定，可追溯至ILAC MRA 國際實驗室認證聯盟相互承認協議成員之認可實驗室，或各國家計量標準機構(NMI)，或國際度量衡委員會相互認可協定之機構(CIPM MRA)，或驗證參考物質(CRMs)。
- ◆ 有TAF認證標誌之報告係全國認證基金會(TAF)之認證範圍；無TAF認證標誌之報告亦符合本實驗室標準校正作業程序及ISO/IEC 17025之規定。
- ◇ SGS Taiwan Ltd hereby declares that traceability follows ILAC P10(ILAC policy on the traceability of measurement results). All standards are directly traceable to TAF recognized lab (members of the ILAC MRA) or to National Metrology Institutes (NMI) or to other international standards (members of the CIPM MRA) or certified reference materials(CRMs). This report with "TAF" accredited symbol indicates the quality system conforms to TAF; Without "TAF" accredited symbol, the report also complies with the lab's standard calibration operating procedures and ISO/IEC 17025 requirements.
- ◆ 本校正報告僅對上述儀器之校正項目有效，本實驗室依ISO/IEC 17025規定不做校正週期及允收水準之判定。
- ◇ This calibration report is valid only to the items been calibrated. According to ISO / IEC 17025, SGS will not provide the determination for Calibration interval and acceptable level for this instrument.
- ◆ 本校正報告部份複製及影本無效。
- ◇ To reproduce or copy calibration report in partial is not allowed.


 報告簽署人

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/terms-and-conditions.aspx> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/terms-and-conditions/terms-and-conditions.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.



溫度部份：

器示值 (°C)	標準值 (°C)	器差值 (°C)
1.7	2.1	-0.4
7.5	8.0	-0.5
-19.2	-19.9	0.7

校正說明：

1. 器差值 = 器示值 - 標準值
2. 器差% = $[(\text{器示值} - \text{標準值}) \div \text{標準值}] * 100$
3. 器示值係指送校正件所顯示或設定之值
4. 標準值係指工作標準件之輸出值或顯示值
5. 校正能力係以約 95 %信賴水準, $k = 2$ 之擴充不確定度表示
6. 擴充不確定度: 1.0 °C
7. 上述擴充不確定度已包含待校件之不確定度評估結果

-- THE END --